## Department of Forensic Science Breath Alcohol Section Instrument Location History 1/1/1990 To 9/26/2013

NOL

Instrument Ser	ial Number: 01060	As of 26-Sep-13	
Start Date	End Date	Agency Name	
25-Sep-13		DFS Central Lab	
14-Mar-13	25-Sep-13	Fairfax County PD	
05-Nov-12	14-Mar-13	DFS Central Lab	
24-May-12	05-Nov-12	Virginia Beach PD	
27-Mar-12	24-May-12	DFS Central Lab	
22-Nov-11	27-Mar-12	Richmond PD	
13-Sep-10	22-Nov-11	DFS Central Lab	
04-Jun-10	13-Sep-10	Intoximeters, Inc.	
12-May-09	04-Jun-10	DFS Central Lab	
18-Mar-09	12-May-09	Norfolk PD	
04-Aug-08	18-Mar-09	DFS Central Lab	

9/25/2013

## INTOX EC/IR II

## Quality Assurance Worksheet

Instrument Serial Number 010603

010603 Worksheet Start Date

Location Fairfax ADC

Address 10520 Judicial Drive, Fairfax VA, 22030

DFS Technician Timmy Neece, Jr License No. 25387

☐ Laboratory ☐ On-Site

Site Specifications Yes No

1 Secured Environment:

Location must not be accessible to the general public.

2 Environment free from contaminants

3 Dedicated telephone line: 

√ 

√

The phone line should be used solely by the instrument.

4 Adequate Ventilation:

Enough space must be provided for the instrument and the printer to allow for adequate ventilation.

5 Stable temperature controls:

The instrument should not be exposed to direct sources of temperature variation.

6 115VAC power source: 

□ □ □

Whenever possible instrument power circuit should not be shared with refrigerators, copy machines or other high amperage devices.

7 Room Condition:

The site should be free of excessive moisture, heat and dust.

Old serial no. New serial no. Supplies

2000104201 Mouthpieces

E-Seek
Keyboard
Printer

2000104201
Mouthpieces
Certificates of Analysis
8ETJXH62423
Operator Worksheet

UPS 10+123 Operator Worksheet 8B0740R19788 Other:

Maintenance (if any)

Comments (if any)

Unable to perform measurement assurance check or instrument test due to instrument being disabled.

Instrument removed to DFS Central.

Intox EC/IR II QA Worksheet Issued by Calibration and Training Program Manager Issue Date: 4-March-2013 DFS Document 250-F113

Revision 6

Page 1 of 2

Instrument Serial Number	010603	Certification Da		te programme to the second sec
Instrument Barometer (mm HG)				meter (mm HG)
Reference Barometer(RB)Serial #			RB Calibration I	Due
Standard (sea level) 0.250 Precision 0	<del>-</del>	minimum 0.000 sample min 0.000	maximum 0.000 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea level) 0.159 Precision 0	· · · · · · · · · · · · · · · · · · ·	minimum 0.000 sample min 0.000	maximum 0.000 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea level) 0.080 Precision 0		minimum -0.003 sample min 0.000	maximum 0.003 sample max 0.000	Sample 1 Sample 2 Sample 3
Standard (sea level) 0.020 Precision 0		minimum -0.003 sample min 0.000	maximum 0.003 sample max 0.000	Sample 1 Sample 2 Sample 3
all measurements are Estimation of Uncertainty of Measi		ability records a	are located withir	n the Breath Alcohol Section
Г Calibrated	Certified		Replaced dry	gas standard (+O-ring)
Dry gas standard Lot N	lo. (with tank no.)		AG210901-09	
Measurement Assurance Che				Instrument Test) s satisfactory condition of the item)
Standard (sea level) 0.100 Precision 0		minimum -0.003 sample min 0.000	maximum 0.003 sample max 0.000	Sample 1 Sample 2 Sample 3
DFS Technician	9:Ny		[	Date <u>9-26-73</u>
Issuing Analyst	Nelissa	Kenned	<u>(1</u>	Date (0/3//2

Intox EC/IR II QA Worksheet
Issued by Calibration and Training Program Manager
Issue Date: 4-March-2013

DFS Document 250-F113

Revision 6 Page 2 of 2



## **IntoxNet MIS Report**

System Communication

Systems Communications Report

Report Generated: 26 Sep 2013 @ 08:05

NISE

Call ID	Site Name (IntoxNet ID)	Call Date a	nd Time	Duration	Completed?	If Not Completed, Reason
00069648	Ffx ADC 010603 N	25 Sep	13:27	95	Т	